

Product:	Anti-Buprenorphine/Norbuprenorphine Antibody  Not available  Norbuprenorphine-BTG					
Synonyms:						
Immunogen:						
	1 mg PAS9857	Host Species:	Ovis aries (Sheep)	Isotype:	IgG	
Pack sizes: Cat. No.	10 mg PAS9841	Host Breed:	Texel	Format:	Clear liquid	
		Clonality:	Polyclonal	pH:	7.4	
IgG Concentration: Lot dependent. Determined @ 280nm.		Buffer: 20mM Phosphate	Buffer: 20mM Phosphate, 150mM Sodium Chloride.			
Recommended Working Concentration*:  0.625 µg/mL  'The application notes include recommended starting dilutions; optimal dilutions/concentrations should be determined by the end user.		Preservative: 0.09% Sodium Az	Preservative: 0.09% Sodium Azide.			
Assessment Method:		Species Reactiv	Species Reactivity:			
Competitive ELISA.		N/A	N/A			
Method of Purification: Salt fractionation.		For long term sto Avoid repeated fr	Storage:  Can be stored for up to 3 months at +2°C - +8°C.  For long term storage, aliquot and store at ≤-20 °C.  Avoid repeated freeze/thaw cycles.  Product should be protected from light exposure.			
Recommende	d Applications:	I.				
The antibody is	suitable for the development of	immunoassays or imm	nunoaffinity purification	columns.		
Sensitivity:  10 ng/mL Buprenorphine HCl produces 88% inhibition in a competitive ELISA employing Buprenorphine/Norbuprenorphine polyclonal antibody.		10 ng/mL Norbup	Target Specificity*:  10 ng/mL Norbuprenorphine produces 85% inhibition in a competitive ELISA employing polyclonal antibody at 1.25 μg/mL.			
		* Cross reactivity profile may	* Cross reactivity profile may vary with tracer used.			
Related Produ	icts:	•				

Buprenorphine/Norbuprenorphine-HRP, HRP9403, Kit size: 0.5 mL.

## Notes & Precautions:

Antibody can be affinity purified on request. This product as supplied is intended for research applications only, not for use in therapeutic or diagnostic applications without the expressed written authorization of Randox BioReagents. A safety data sheet (SDS) can be supplied upon request. Vial should be centrifuged briefly before opening to ensure all material is removed from the vial cap.

28-Apr-22

Information correct at time of going to print

